


	CERTIFICATE OF ANALYSES Certificat d'analyses		Diffusion : Réaliser Produits
	Approuvé par le pilote : NB	Vérfié par Responsable Qualité : LR	MAJ : 22/08/2023 Page : 1/1

Designation :	Id-Fungi Plates	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	19/02/2024		
Expiration date :	03/09/2024		36900
Control date :	20/02/2024		
PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS	
Appearance	Translucent Beige	CONFORM	
pH	4,5 to 6,0 (indicative value)	pH=5,3	
WIDTH (mm)	>3mm	CONFORM	
MICROBIOLOGICAL STATE			
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM	
MICROBIOLOGICAL ACTIVITY			
Inoculum < 100 cfu Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM	
Inoculum > 100 cfu Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM	
Inoculum : NA Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM	
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM	
CONFORM			
AQ CONCLUSION	BATCH ACCEPTED		

Date : 12 / 03 / 2024.


 CONIDIA CONIPHY
 2 rue en Chuel
 PA en Chuel
 69650 Quincieux – France

Responsable Qualité


 NB