



Designation :	Id-Fungi Plates	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	29/01/2024		
Expiration date :	13/08/2024		36728
Control date :	30/01/2024		
PHYSICAL-CHEMICAL TESTS		SPECIFICATIONS	RESULTS
Appearance	Translucent Beige		CONFORM
pH	4,5 to 6,0 (indicative value)		pH=5,2
WIDTH (mm)	>3mm		CONFORM
MICROBIOLOGICAL STATE			
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C		CONFORM
MICROBIOLOGICAL ACTIVITY			
Inoculum < 100 cfu Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth		CONFORM
Inoculum > 100 cfu Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition		CONFORM
Inoculum : NA Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth		CONFORM
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth		CONFORM
		CONFORM	
AQ CONCLUSION			BATCH ACCEPTED

Date : 13/02/2024

Responsable Qualité


 CONIDIA CONIPHY
 2 rue en Chuel
 PA en Chuel
 69650 Quincieux – France