

Designation :	<b>Id-Fungi Plates</b>	BATCH N° :
Presentation :	Petri dishes 90 mm	
Production date :	29/04/2024	
Expiration date :	12/11/2024	<b>37579</b>
Control date :	30/04/2024	

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance	Translucent Beige	CONFORM
pH	4,5 to 6,0 (indicative value)	pH=5,1
WIDTH (mm)	>3mm	CONFORM
MICROBIOLOGICAL STATE		
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY		
<b>Inoculum &lt; 100 cfu</b> Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM
<b>Inoculum &gt; 100 cfu</b> Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM
<b>Inoculum : NA</b> Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM
<b>CONFORM</b>		
<b>AQ CONCLUSION</b>	<b>BATCH ACCEPTED</b>	

Date :

22 / 05 / 2024

Responsable Qualité



CONIDIA CONIPHY  
 2 rue en Chuel  
 PA en Chuel  
 69650 Quincieux – France



*[Handwritten signature]*