

Designation :	<b>Id-Fungi Plates</b>	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	14/05/2024		
Expiration date :	27/11/2024		<b>37626</b>
Control date :	17/05/2024		
PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS	
Appearance	Translucent Beige	CONFORM	
pH	4,5 to 6,0 (indicative value)	pH=5,3	
WIDTH (mm)	>3mm	CONFORM	
MICROBIOLOGICAL STATE			
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM	
MICROBIOLOGICAL ACTIVITY			
<b>Inoculum &lt; 100 cfu</b> Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM	
<b>Inoculum &gt; 100 cfu</b> Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM	
<b>Inoculum : NA</b> Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM	
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM	
<b>CONFORM</b>			
<b>AQ CONCLUSION</b>	<b>BATCH ACCEPTED</b>		

Date :

04/06/2024

Responsable Qualité



CONIDIA CONIPHY  
 2 rue en Chuel  
 PA en Chuel  
 69650 Quincieux – France

