

Designation :	Id-Fungi Plates	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	28/06/2024		
Expiration date :	11/01/2025		37999
Control date :	01/07/2024		
PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS	
Appearance	Translucent Beige	CONFORM	
pH	4,5 to 6,0 (indicative value)	pH=5,3	
WIDTH (mm)	>3mm	CONFORM	
MICROBIOLOGICAL STATE			
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM	
MICROBIOLOGICAL ACTIVITY			
Inoculum < 100 cfu Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM	
Inoculum > 100 cfu Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM	
Inoculum : NA Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM	
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM	
CONFORM			
AQ CONCLUSION	BATCH ACCEPTED		

Date :

23 / 07 / 24

Responsable Qualité:



CONIDIA CONIPHY
 2 rue en Chuel
 PA en Chuel
 69650 Quincieux – France