

Approuvé parle pilote : CD      Vérifié par Responsable Qualité : NB

Designation : **Id-Fungi Plates**      BATCH N° :  
 Presentation : Petri dishes 90 mm  
 Production date : 02/08/2024  
 Expiration date : 15/02/2025      **38337**  
 Control date : 05/08/2024

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance pH WIDTH (mm)	Translucent Beige 4,5 to 6,0 (indicative value) >3mm	CONFORM pH=5,5 CONFORM
MICROBIOLOGICAL STATE		
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY		
<b>Inoculum &lt; 100 cfu</b> Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM
<b>Inoculum &gt; 100 cfu</b> Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM
<b>Inoculum : NA</b> Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM

**CONFORM**

**AQ CONCLUSION**      **BATCH ACCEPTED**

Date : 22 / 08 / 2024

Responsable Qualité




CONIDIA CONIPHY  
2 rue en Chuel  
PA en Chuel  
69650 Quincieux – France

