



FORMULAIRE:

CERTIFICATE OF ANALYSES / CERTIFICAT D'ANALYSES

Diffusion : Réaliser Produits

Approuvé parle pilote : CD

Vérfifié par Responsable Qualité : NB

MAJ : 11/09/2024

Page : 1/1

NOM Produit :	Id-Fungi Plates	BATCH N° :
Presentation :	Petri dishes 90 mm	
Production date :	03/09/2024	
Expiration date :	19/03/2025	<b>38424</b>
Control date :	05/09/2024	

PHYSICAL-CHIMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance pH	Beige indicative value	CONFORM pH=5.3
MICROBIOLOGICAL STATE	SPECIFICATIONS	RESULTS
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY	SPECIFICATIONS	RESULTS
<b>Inoculum &lt; 100 CFU</b>  <i>Candida albicans</i> ATCC 10231 / WDCM 00054	After 48H-72H at 20-25°C GROWTH	CONFORM
<b>Inoculum &gt; 100 CFU</b>  <i>Escherichia coli</i> ATCC 8739 / WDCM 00012	After 48H at 20-25°C INHIBITION	CONFORM
<b>Inoculum : N/A</b>  <i>Aspergillus brasiliensis</i> ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C GROWTH	CONFORM
<i>Trichophyton rubrum</i> IP 1464.83	5 to 10 days at 20-25°C GROWTH	CONFORM

<b>CONCLUSION</b>	<b>CONFORM</b>	<b>NOT CONFORM</b>
CQ Name: <u>VACHER</u>	<input checked="" type="checkbox"/> BATCH ACCEPTED <input type="checkbox"/> BATCH NOT ACCEPTED	

Date : 19/09/2024

Quality Manager:

CONIDIA CONIPHY  
2 rue en Chuel  
PA en Chuel  
69650 Quincieux – France