



FORMULAIRE:

CERTIFICATE OF ANALYSES / CERTIFICAT D'ANALYSES

Diffusion : Réaliser Produits

Approuvé parle pilote : CD

Vérfifié par Responsable Qualité : NB

MAJ : 11/09/2024

Page : 1/1

Product name : Id-Fungi Plates BATCH N° :
 Presentation : Petri dishes 90 mm
 Production date : 14/10/2024
 Expiration date : 29/04/2025 **38743**
 Control date : 17/10/2024

| PHYSICAL-CHEMICAL TESTS | SPECIFICATIONS | RESULTS |
|-------------------------|---------------------------|-------------------|
| Appearance pH | Beige indicative value | CONFORM pH=5.3 |

| MICROBIOLOGICAL STATE | SPECIFICATIONS | RESULTS |
|---|--|---------|
| Direct incubation before and after the membrane set up | Sterile after 7 days at 20-25°C and 30-35°C | CONFORM |

| MICROBIOLOGICAL ACTIVITY | SPECIFICATIONS | RESULTS |
|---|------------------------------------|---------|
| Inoculum < 100 CFU <i>Candida albicans</i> ATCC 10231 / WDCM 00054 | After 48H-72H at 20-25°C GROWTH | CONFORM |
| Inoculum > 100 CFU <i>Escherichia coli</i> ATCC 8739 / WDCM 00012 | After 48H at 20-25°C INHIBITION | CONFORM |
| Inoculum : N/A <i>Aspergillus brasiliensis</i> ATCC 16404 / WDCM 00053 | Until 5 days at 20-25°C GROWTH | CONFORM |
| <i>Trichophyton rubrum</i> IP 1464.83 | 5 to 10 days at 20-25°C GROWTH | CONFORM |

| CONCLUSION | CONFORM | NOT CONFORM |
|--------------------|---|-------------|
| CQ Name: <u>CD</u> | <input checked="" type="checkbox"/> BATCH ACCEPTED <input type="checkbox"/> BATCH NOT ACCEPTED | |

Date : 31/10/24

Quality Manager:

CONIDIA CONIPHY
 2 rue en Chuel
 PA en Chuel
 69650 Quincieux – France

