

Designation :	Id-Fungi Plates Plus	BATCH N° :
Presentation :	Petri dishes 90 mm	
Production date :	02/05/2024	
Expiration date :	15/11/2024	37578
Control date :	03/05/2024	

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance	Translucent Beige	CONFORM
pH	4,5 to 6,0 (indicative value)	pH=5,3
WIDTH (mm)	>3mm	CONFORM
MICROBIOLOGICAL STATE		
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY		
Inoculum > 100 cfu Escherichia coli ATCC 8739 / WDCM 00012	<i>After 48h at 20-25°C</i> Inhibition	CONFORM
Inoculum : NA Aspergillus brasiliensis ATCC 16404 / WDCM 00053	<i>Until 5 days at 20-25°C</i> Inhibition	CONFORM
Trichophyton rubrum IP 1464.83	<i>5 to 10 days at 20-25°C</i> Growth	CONFORM

AQ CONCLUSION	CONFORM
	BATCH ACCEPTED

Date :

22 / 05 / 2024

Responsable Qualité



CONIDIA CONIPHY
2 rue en Chuel
PA en Chuel
69650 Quincieux – France



[Handwritten signature]