



FORMULAIRE:

**CERTIFICATE OF ANALYSES / CERTIFICAT D'ANALYSES**

Diffusion : Réaliser  
Produits

Approuvé par le pilote : CD

Vérifié par Responsable Qualité : NB

MAJ : 22/11/2024


Page : 1/1

NOM Produit : Id-Fungi Plates BATCH N° :  
 Presentation : Petri dishes 90 mm  
 Production date : 12/11/2024  
 Expiration date : 28/05/2025 **38902**  
 Control date : 13/11/2024


PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance pH	Beige indicative value	CONFORM pH=5.3


MICROBIOLOGICAL STATE	SPECIFICATIONS	RESULTS
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM

MICROBIOLOGICAL ACTIVITY	SPECIFICATIONS	RESULTS
<b>Inoculum &lt; 100 CFU</b>  <i>Candida albicans</i> ATCC 10231 / WDCM 00054	After 48H-72H at 20-25°C GROWTH	CONFORM
<b>Inoculum &gt; 100 CFU</b>  <i>Escherichia coli</i> ATCC 8739 / WDCM 00012	After 48H at 20-25°C INHIBITION	CONFORM
<b>Inoculum : N/A</b>  <i>Aspergillus brasiliensis</i> ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C GROWTH	CONFORM
<i>Trichophyton rubrum</i> IP 1464.83	5 to 10 days at 20-25°C GROWTH	CONFORM

CONCLUSION	CONFORM	NOT CONFORM
AQ Name: 	<input checked="" type="checkbox"/> BATCH ACCEPTED <input type="checkbox"/> BATCH NOT ACCEPTED	

Date : 02/12/24

Quality Manager: 

 CONIDIA CONIPHY  
2 rue en Chuel  
PA en Chuel  
69650 Quincieux – France